

LOT 03231C

NAME	_____	DATE	_____
ID#	_____	TECH	_____
BDT	_____		


EXP: 2012-03


Suggested Cutoff SCO	1500	Suggested Cutoff SCO	5	Suggested Cutoff SCO	5
-------------------------	-------------	-------------------------	----------	-------------------------	----------

BEAD	PROBE	CLASS II ANTIGEN		BACKGROUND MFI		Raw MFI	Adjusted Value 1 (Raw - Background)	Adjusted Value 2 (Adjusted Value 1 / CalcCON)	Relative amount of antigen	Adjusted Value 3 (Adjusted Value 2 / Relative amount of antigen)	Interpretation
		CLII	AG	BKG	MF	MF	AV ₁	CalcCON	AV ₂	AG-O	AV ₃
8	108	DPB1*01:01/DPA1*01:03		714					0.696		
9	109	DPB1*01:01/DPA1*02:01		588					0.519		
77	110	DPB1*01:01/DPA1*02:02		609					0.644		
78	111	DPB1*01:01/DPA1*03:01		536					0.671		
12	112	DPB1*02:01/DPA1*01:03		754					0.572		
79	113	DPB1*03:01/DPA1*01:03		478					0.635		
14	114	DPB1*04:01/DPA1*01:03		809					0.421		
15	115	DPB1*04:01/DPA1*02:01		605					0.457		
16	116	DPB1*04:01/DPA1*02:02		602					0.398		
17	117	DPB1*04:01/DPA1*03:01		707					0.418		
80	118	DPB1*04:01/DPA1*04:01		635					0.672		
19	119	DPB1*04:02/DPA1*01:03		827					0.374		
20	120	DPB1*04:02/DPA1*03:01		796					0.468		
81	121	DPB1*05:01/DPA1*02:01		637					0.664		
82	122	DPB1*05:01/DPA1*02:02		757					0.553		
98	123	DPB1*05:01/DPA1*03:01		609					0.568		
24	124	DPB1*13:01/DPA1*02:01		788					0.338		
25	125	DPB1*13:01/DPA1*04:01		703					0.537		
26	126	DPB1*14:01/DPA1*02:01		615					0.558		
27	127	DPB1*17:01/DPA1*02:01		690					0.676		
83	128	DPB1*18:01/DPA1*01:03		895					0.577		
84	129	DPB1*19:01/DPA1*02:01		621					0.464		
85	130	DPB1*28:01/DPA1*02:02		739					0.487		
31	131	DQB1*02:02/DQA1*02:01		1092					0.331		
32	132	DQB1*02:02/DQA1*03:02		904					0.593		
33	133	DQB1*02:02/DQA1*05:01		956					0.704		
34	134	DQB1*03:01/DQA1*03:02		614					0.707		
86	135	DQB1*03:01/DQA1*05:01		731					0.563		
36	136	DQB1*03:02/DQA1*02:01		592					0.606		
37	137	DQB1*03:02/DQA1*03:02		544					0.627		
38	138	DQB1*03:03/DQA1*04:01		614					0.596		
39	139	DQB1*03:03/DQA1*06:01		398					0.891		
87	140	DQB1*04:01/DQA1*04:01		544					0.651		
41	141	DQB1*04:02/DQA1*04:01		561					0.710		
42	142	DQB1*04:02/DQA1*06:01		380					0.757		
91	143	DQB1*05:01/DQA1*01:01		514					0.860		
89	144	DQB1*05:01/DQA1*01:02		519					0.705		
45	145	DQB1*05:03/DQA1*01:04		818					0.510		
46	146	DQB1*05:03/DQA1*02:01		454					0.799		
47	147	DQB1*06:01/DQA1*01:04		435					0.865		
99	148	DQB1*06:01/DQA1*02:01		1030					0.614		
88	149	DQB1*06:01/DQA1*05:01		503					0.918		
50	150	DQB1*06:03/DQA1*01:03		626					0.806		
51	151	DRB1*01:01		807					0.481		

BEAD	PROBE	CLASS II ANTIGEN		BACKGROUND MFI		Raw MFI	Adjusted Value 1 (Raw - Background)	CalcCON	Adjusted Value 2 (Adjusted Value 1/ CalcCON)	Relative amount of antigen	Adjusted Value 3 (Adjusted Value 2/ Relative amount of antigen)	Interpretation
		CLII	AG	BKG	MF	MF	AV ₁		AV ₂	AG-O	AV ₃	INTRP
92	152	DRB1*01:02		554					0.601			
53	153	DRB1*01:03		678					0.529			
54	154	DRB1*04:01		766					0.585			
55	155	DRB1*04:02		717					0.562			
56	156	DRB1*04:05		598					0.526			
94	157	DRB1*07:01		595					0.654			
58	158	DRB1*08:01		490					0.632			
93	159	DRB1*09:01		713					0.635			
97	160	DRB1*10:01		934					0.522			
61	161	DRB1*11:01		624					0.676			
62	162	DRB1*12:01		605					0.545			
63	163	DRB1*13:01		606					0.644			
64	164	DRB1*13:03		546					0.677			
100	165	DRB1*14:01		619					0.702			
66	166	DRB1*15:01		508					0.565			
67	167	DRB1*15:02		800					0.848			
68	168	DRB1*16:01		656					0.738			
69	169	DRB1*03:01		752					0.602			
70	170	DRB1*03:03		625					0.683			
71	171	DRB3*01:01		681					1.000			
95	172	DRB3*02:02		714					0.774			
73	173	DRB3*03:01		605					0.780			
96	174	DRB4*01:01		498					0.619			
75	175	DRB5*01:01		741					0.872			
76	176	DRB5*02:02		719					0.683			
2	CON1	CON1										
3	CON2	CON2										
4	CON3	CON3										
6	77	POS CONTROL	IgG									

LSA Class II Positive Control - EXPECTED RESULT

CON1	211-396 MFI
CON2	132-485 MFI
CON3	158-441 MFI

Class II Positive Control Sera, Lot 10230D

