



# AccuProbe

## Blastomyces Dermatitidis Culture Identification Test

### Intended Use

The ACCUPROBE Blastomyces Dermatitidis Culture Identification Test is a rapid DNA probe test which utilizes the technique of nucleic acid hybridization for the identification of *Blastomyces dermatitidis* isolated from culture.

### Summary and Explanation of the Test

*Blastomyces dermatitidis* (*B. dermatitidis*) is the causative agent of the blastomycosis (North American blastomycosis or Gilchrist's disease), a systemic fungal disease usually occurring as a chronic granulomatous infection of the skin or internal organs. Infection is thought to occur through inhalation of spore-laden dust. From these ports of entry, the organism can spread to the bones, kidneys, lymph nodes, skin, and central nervous system (2, 4, 10, 12).

Blastomycosis is not transferable from person to person or animal to person. It occurs sporadically; individuals engaged in outdoor and agricultural professions are more likely to contract blastomycosis. Most cases occur in the central and southeastern United States and Canada. Indigenous cases have also been reported in South America, Central America, the Middle East, and Africa (12). *B. dermatitidis* is a thermally dimorphic fungus which, under different environmental conditions, may exist in either the mold or yeast phase. Growth may occur as a filamentous mold at room temperature on media such as Sabouraud Dextrose Agar or in the large broad-based budding yeast form on media such as Brain Heart Infusion Agar at 35° to 37°C (6, 7, 10).

Laboratory identification methods used to identify *B. dermatitidis* include culturing clinical specimens on nutrient media, colony morphology, and characteristic microscopic sporulation patterns, filamentous to yeast phase conversion and exoantigen tests. Conventional laboratory identification methods begin with culture of the clinical specimen on specialized fungal media. The time required for growth of *B. dermatitidis* to a visible colony varies from a few days to over two months. The colony morphology of this organism is also quite variable, ranging from flat, glabrous colonies to fluffy-white or brownish colonies (6, 10). Additional growth is needed before the characteristic microscopic sporulation pattern of short conidiophores bearing pyriform conidia may be seen. Many species of fungi other than *B. dermatitidis* may produce similar colony and sporulation characteristics. These species include *Histoplasma capsulatum* and *Chrysosporium* spp. (8). Therefore additional testing is needed to definitively identify this organism. One method is to convert the room temperature mycelial phase growth to the yeast phase by sub-culturing onto Brain Heart Infusion, cottonseed or blood agar and incubating at 35° to 37°C. Within a few days or weeks, the characteristic large, thick-walled, broad-based yeast cells may be visible upon microscopic examination. Other confirmatory tests based on exoantigen extraction have been described (6, 10). These tests may take from three to five days or longer to perform.

The ACCUPROBE Blastomyces Dermatitidis Culture Identification Test identifies both phases regardless of sporulation in less than an hour of sample preparation based on identification of specific ribosomal RNA sequences that are unique to *B. dermatitidis*.

### Principles of the Procedure

Nucleic acid hybridization tests are based on the ability of complementary nucleic acid strands to specifically align and associate to form stable double-stranded complexes (5). The ACCUPROBE SYSTEM uses a single-stranded DNA probe with a chemiluminescent label that is complementary to the ribosomal RNA of the target organism. After the ribosomal RNA is released from the organism, the labeled DNA probe combines with the target organism's ribosomal RNA to form a stable DNA:RNA hybrid. The Selection Reagent allows for the differentiation of non-hybridized and hybridized probe. The labeled DNA:RNA hybrids are measured in a LEADER luminometer. A positive result is a luminometer reading equal to or greater than the cut-off. A value less than this cut-off is a negative result.

### Reagents

Reagents for the ACCUPROBE Blastomyces Dermatitidis Culture Identification Test are provided in three separate reagent kits:

#### ACCUPROBE Blastomyces Dermatitidis Probe Kit

Probe Reagent (10 x 2 tubes)

Blastomyces dermatitidis

Lysing Reagent (1 x 20 tubes)

Glass beads and buffer

#### ACCUPROBE Culture Identification Reagent Kit

Reagent 1 (Lysis Reagent) 1 x 10 mL

buffered solution containing 0.04% sodium azide

Reagent 2 (Hybridization Buffer) 1 x 10 mL

buffered solution

Reagent 3 (Selection Reagent) 1 x 60 mL

buffered solution

#### GEN-PROBE Detection Reagent Kit

Detection Reagent I 1 x 240 mL

0.1% hydrogen peroxide in 0.001 N nitric acid

Detection Reagent II 1 x 240 mL

1 N sodium hydroxide

### Warnings and Precautions

- A. For *in vitro* diagnostic use.
- B. Use universal precautions when performing this assay (1).
- C. Use only for the identification of *B. dermatitidis* isolated from culture.
- D. Use only supplied or specified disposable laboratory ware.
- E. Culture handling and all procedural steps through the heat inactivation step should be performed in a Class II Biological Safety Cabinet.
- F. Reagents in this kit contain sodium azide, which may react with lead or copper plumbing to form potentially explosive metal azides. Upon disposal of these reagents, always dilute the material with a large volume of water to prevent azide buildup in the plumbing.
- G. Avoid contact of Detection Reagents I and II with skin and mucous membranes. Wash with water if these reagents come into contact with skin. If spills of these reagents occur, dilute with water before wiping dry.

### Storage and Handling Requirements

Probe Reagent Tubes must be stored in the foil pouches at 2° to 8°C. The Probe Reagent Tubes are stable in the unopened pouches until the expiration date indicated. Once opened, the pouch should be resealed and the tubes should be used within two months and prior to the expiration date.

Other reagents used in the ACCUPROBE Blastomyces Dermatitidis Culture Identification Test may be stored between 2° and 25°C and are stable until the expiration date indicated.

Do not freeze the reagents.

## Sample Collection and Preparation

The ACCUPROBE *Blastomyces Dermatitidis* Culture Identification Test kit is designed to determine the identity of *Blastomyces dermatitidis* isolated from culture. Yeast colonies should be no more than one week old from the time growth is first observed. Filamentous colonies should be no more than one month old from the time growth is first observed. Sporulation is not necessary.

A. Solid Media Method: Growth from appropriate solid media such as Sabouraud Dextrose, Brain Heart Infusion, Mycobiotic (Mycosel), Inhibitory Mold Agar, Cottonseed Agar, Yeast Nitrogen Base Agar or 5% Sheep Blood Agar suggestive of *B. dermatitidis* may be tested.

1. Growth can be removed with a 1 µL disposable plastic loop, a wire loop, a disposable plastic needle, or an applicator stick. Swabs should not be used due to the small volume of liquid in which the cells are subsequently resuspended.
2. Avoid taking large amounts of the solid media with the cells.
3. A 1 - 2 mm<sup>2</sup> size sample of yeast or filamentous growth is recommended for the test.
4. The operator may elect to inoculate another culture plate at this time to confirm the purity of the isolate.

B. Broth Culture Method: Growth in Brain Heart Infusion Broth with turbidity equivalent to or greater than a McFarland 1 Nephelometer Standard may be tested with the ACCUPROBE *Blastomyces Dermatitidis* Culture Identification Test. Pipette a 100 µL sample from the well-mixed broth suspension into the Lysing Reagent Tube as described below.

## Materials Provided

The ACCUPROBE *Blastomyces Dermatitidis* Culture Identification Test:

<b>Cat. No. 102890</b>	<b>20 Tests</b>
Probe Tubes	10 x 2 tubes
Lysing Reagent	1 x 20 tubes

## Materials Required but not Provided

One (1) µL plastic sterile inoculating loops, wire loops, plastic needles, or applicator sticks for selecting colonies.

Control culture strains

Water bath or heating block (60° ± 1°C)

Water bath or heating block (95° ± 5°C)

Micropipettes (100 µL)

Re-pipettor (100 µL, 300 µL)

Vortex Mixer

## Materials Available from Gen-Probe

LEADER Luminometer

GEN-PROBE Ultrasonic Water Bath (Cat. No. 901104 or equivalent)

ACCUPROBE Culture Identification Reagent Kit (Cat. No. 102800)

GEN-PROBE Detection Reagent Kit (Cat. No. 201791)

GEN-PROBE Heating Block (Cat. No. Single Block 302902; Triple Block 302903)

GEN-PROBE Ultrasonic Water Bath Rack (Cat. No. 901393)

## Test Procedure

### A. Equipment Preparation

1. For optimal transfer of sonic energy, water must be thoroughly degassed according to the following procedure:
  - a. Add enough hot water to fill the ultrasonic water bath to within 1/2 inch from the top of the tank.
  - b. Run the ultrasonic water bath for 15 minutes to thoroughly degas the water.
2. Adjust one heating block or water bath to 60° ± 1°C and another heating block or water bath to 95° ± 5°C.
3. Prepare the LEADER luminometer for operation. Make sure there is sufficient volume of Detection Reagents I and II to complete the tests.

### B. Controls

Positive and negative control strains should be tested routinely in each laboratory according to local regulations. A culture of *B. dermatitidis* (e.g., American Type Culture Collection, ATCC # 60916) may be used as the positive control while a culture of *Histoplasma capsulatum* (e.g., ATCC # 38904) may be used as the negative control.

### C. Sample Preparation

1. Label a sufficient number of Lysing Reagent Tubes to test the culture isolates and/or controls. Remove and retain the caps.
2. Pipette 100 µL of Reagent 1 (Lysis Reagent) and 100 µL of Reagent 2 (Hybridization Buffer) into the Lysing Reagent Tubes. If broth cultures are to be tested, do not add Reagent 1 to the Lysing Reagent Tubes.
3. Transfer the sample from the solid media or 100 µL of a well-mixed broth culture into labeled Lysing Reagent Tubes as described in the *Sample Collection and Preparation* section. Twirl the loop, needle or stick in the Reagent 1 and Reagent 2 diluent mixture to remove the cells if testing growth from solid medium.
4. Recap the Lysing Reagent Tubes and briefly **vortex**.

### D. Sample Lysis

1. Push the Lysing Reagent Tubes through the ultrasonic water bath rack so that the reaction mixture in the bottom of the tube is submerged but the caps are above the water. Place the ultrasonic water bath rack on the ultrasonic water bath. **Do not allow the tubes to touch the bottom or sides of the ultrasonic water bath.**
2. Sonicate for 15 minutes.
3. Place the Lysing Reagent Tubes containing the sonicated organisms in a heating block or water bath for 10 minutes at 95° ± 5°C.
4. Carefully remove the Lysing Reagent Tubes from the heating block or water bath.

### E. Hybridization

1. Open the foil pouch by cutting evenly across the top of the pouch. Remove enough Probe Reagent Tubes to test the culture isolates and/or controls. Reseal the pouch by folding the opened edge over several times and securing the fold with adhesive tape or a clip. **Leave the desiccant pillow in the pouch.**
2. Label a sufficient number of Probe Reagent Tubes to test the culture isolates and/or controls. Remove and retain the caps.
3. Pipette 100 µL of the lysed specimens from the Lysing Reagent Tubes into the corresponding Probe Reagent Tubes.

- Recap the Probe Reagent Tubes and incubate for 15 minutes at 60° ± 1°C in a water bath or heating block.

F. Selection

- Remove the Probe Reagent Tubes from the water bath. Remove and retain the caps. Pipette 300 µL of Reagent 3 (Selection Reagent) into each tube. If a heating block is used, Reagent 3 may be pipetted directly into the Probe Reagent Tubes. Recap the tubes and **vortex** the tubes to mix completely.
- Incubate the Probe Reagent Tubes for 5 minutes at 60° ± 1°C in a water bath or heating block.
- Remove the Probe Reagent Tubes from the water bath or heating block and leave them at room temperature for at least 5 minutes. Remove and discard the caps. Read the results in the luminometer within 1 hour.

G. Detection

- Select the appropriate protocol from the menu of the luminometer software.
- Using a damp tissue or paper towel, wipe each tube to ensure that no residue is present on the outside of the tube and insert the tube into the luminometer according to the instrument directions.
- When the analysis is complete, remove the last tube(s) from the luminometer.

## Procedural Notes

- Reagents:** Reagent 2 (Hybridization Buffer) may precipitate. Warming and mixing the solution at 35° to 60°C will dissolve the precipitate.
- Temperature:** The Hybridization and Selection reactions are temperature dependent. Therefore, it is imperative that the water bath or heating block is maintained within the specified temperature range.
- Time:** The Hybridization and Selection reactions are time dependent. Hybridize at least 15 minutes but no more than 20 minutes. Incubate the Probe Reagent Tubes during the Selection step for at least 5 minutes but no more than 6 minutes.
- Water Bath:** The level of water in the water bath should be maintained to ensure that the Lysing Reagent Tubes are submerged up to, but not above, the level of the sealing ring. It should also be ensured that the entire liquid reaction volume in the Probe Reagent Tubes is submerged.
- Vortexing:** It is critical to have a homogenous mixture during the Sample Preparation and Selection steps; specifically after the addition of cells to Reagents 1 and 2 and after the addition of Reagent 3.
- Troubleshooting:**
  - Elevated negative control values (*H. capsulatum* ATCC # 38904) greater than 20,000 RLU (Relative Light Units) in the LEADER luminometer or 600 PLU (Photometric Light Units) in the ACCULDR (formerly PAL) luminometer can be caused by insufficient mixing after adding Reagent 3 (Selection Reagent) or by testing mixed cultures. Because mixed cultures can occur, a portion of the growth may be streaked onto an appropriate agar medium and incubated to check for multiple colony types.
  - Low positive control values (*Blastomyces dermatitidis* ATCC #60916) less than 50,000 RLU in the LEADER luminometer or 1,500 PLU in the ACCULDR (formerly PAL) luminometer can be caused by insufficient cell numbers, improper sonication, or by testing mixed or aged cultures. Because mixed cultures can occur, a portion of the growth may be streaked onto an appropriate agar medium and incubated to check for multiple colony types.

## Results

### A. Interpretation of Results

The results of the ACCUPROBE Blastomyces Dermatitidis Culture Identification Test are based on the following cut-off values. Samples producing signals greater than or equal to these cut-off values are considered positive. Signals less than these cut-off values are considered negative. Results in repeat ranges should be repeated.

	ACCULDR (formerly PAL)	LEADER
Cut-off value	1,500 PLU	50,000 RLU
Repeat range	1,200-1,499 PLU	40,000-49,999 RLU

### B. Quality Control and Acceptability of Results

Negative control (e.g., *H. capsulatum* ATCC # 38904) and positive control (e.g., *B. dermatitidis* ATCC # 60916) should satisfy the following values:

	ACCULDR (formerly PAL)	LEADER
Negative control	< 600 PLU	< 20,000 RLU
Positive control	> 1,500 PLU	> 50,000 RLU

## Limitations

The ACCUPROBE Blastomyces Dermatitidis Culture Identification Test has been shown to produce a positive result with all *Paracoccidioides brasiliensis* (*P. brasiliensis*) strains that have been tested. The taxonomic classification of *P. brasiliensis* remains controversial. It is considered by some to be in the genus *Blastomyces* (11). *P. brasiliensis* is the causative agent of paracoccidioidomycosis or South American blastomycosis. This fungal disease follows a progressive, granulomatous pattern similar to North American blastomycosis (11). South American blastomycosis is found almost exclusively in South and Central America and is indigenous to South America (9). The occurrence of paracoccidioidomycosis in the Northern Hemisphere is extremely rare. Approximately 5 cases of paracoccidioidomycosis per year are reported in the United States, compared to approximately 1,000 to 2,000 cases of North American blastomycosis per year. (Pers. comm. Dr. L. Ajello, Centers for Disease Control, Atlanta, GA.)

*P. brasiliensis* may be distinguished from *B. dermatitidis* by microscopic examination in the yeast phase. *P. brasiliensis* is differentiated by the growth of multiple buds from a single cell with thinner cell walls, where *B. dermatitidis* is typically a single-budding yeast with thicker cell walls. Although it should be noted that occasionally *P. brasiliensis* may produce single-budding yeast (6, 11).

The ACCUPROBE Blastomyces Dermatitidis Culture Identification Test has been shown to produce a positive result with the mold, *Gymnascella hyalinospora* and with *Emmonsia parva*. These organisms are rarely isolated (3). *Emmonsia parva* rarely occurs in humans.

The ACCUPROBE Blastomyces Dermatitidis Culture Identification Test has been shown to produce a positive result with the mold *Chrysosporium charmichaellii*. The clinical implications of this are not known.

This method has been tested using both fresh growth from agar plates and broth cultures as listed in the *Sample Collection and Preparation* section. The efficacy of this test has not been demonstrated on direct clinical specimens.

Results from the ACCUPROBE Blastomyces Dermatitidis Culture Identification Test should be interpreted in conjunction with other laboratory and clinical data available to the clinician.

## Expected Values

The ACCUPROBE Blastomyces Dermatitidis Culture Identification Test was compared to standard culture, morphological and biochemical identification methods at two sites using a total of 396 isolates. Of these, 108 were *B. dermatitidis* isolates (11 isolates were tested in both the yeast and the mycelial phases), 1 *P. brasiliensis* isolate, and 287 isolates of 68 species from 76 genera, representing a phylogenetic cross-section of organisms. Included among these were

103 isolates of the following dimorphic fungi: 7 *Coccidioides immitis*; 94 *Histoplasma capsulatum*; 2 *Sporothrix schenckii*. Standard culture identification is dependent on microscopic and macroscopic colony morphology, sporulation characteristics, filamentous to yeast phase conversions, and in some cases exoantigen tests. Isolates were either categorized as positive ( $\geq 50,000$  RLU) or negative ( $< 50,000$  RLU). The range of observations for negative isolates was 936 RLU to 32,072 RLU and 66,052 RLU to 644,940 RLU for positive isolates. A comparison of the ACCUPROBE Blastomyces Dermatitidis Culture Identification Test and standard culture identification methods is shown below.

ACCUPROBE Culture	ACCUPROBE / Culture Identification				Sensitivity/ Specificity	Percent Agreement
	Pos	Neg	Pos	Neg		
Site 1	53	1	0	77	100/98.7%	99.2%
Site 2	53	0	2	210	96.4/100%	99.2
<b>Total</b>	<b>106</b>	<b>1</b>	<b>2</b>	<b>287</b>	<b>98.1/99.7%</b>	<b>99.2%</b>

The one positive result at Site 1 was a *P. brasiliensis* isolate. As discussed in the *Limitations* section, the ACCUPROBE Blastomyces Dermatitidis Culture Identification Test does not differentiate between *B. dermatitidis* and *P. brasiliensis* isolates since both have been shown to produce positive results. The sensitivity, specificity, and percent agreement for Site 1, are therefore 100%. The two isolates initially producing a negative result at Site 2 produced positive results upon retesting with the ACCUPROBE Blastomyces Dermatitidis Culture Identification Test. The overall sensitivity, specificity and percent agreement, upon retesting, are therefore 100%.

## Performance Characteristics

### A. Within-Run Precision

The within-run precision of the ACCUPROBE Blastomyces Dermatitidis Culture Identification Test was calculated by assaying two concentrations of ribosomal RNA isolated from *B. dermatitidis* using 10 replicates in a single assay.

Sample	A	B
Number of Replicates	10	10
Mean Response	57,662	135,061
Standard Deviation	5,929	3,694
Coefficient of Variation	10.3%	2.7%

### B. Between-Run Precision

The between-run precision was calculated by assaying the same two concentrations of *B. dermatitidis* ribosomal RNA using single determinations in 10 consecutive runs.

Sample	A	B
Number of Replicates	10	10
Mean Response	59,160	180,121
Standard Deviation	6,656	8,537
Coefficient of Variation	11.3%	4.7%

### C. Specificity

A total of 102 ATCC reference isolates were evaluated using the ACCUPROBE Blastomyces Dermatitidis Culture Identification Test. These isolates represented a total of 71 species from 46 genera. Four isolates of *B. dermatitidis*, 4 isolates of *P. brasiliensis*, 6 isolates of *Histoplasma capsulatum*, and 1 isolate of *Sporothrix schenckii* were tested in the filamentous phase. Eighty-seven isolates of 42 other genera representing a phylogenetic cross-section of organisms were evaluated using the ACCUPROBE Blastomyces Dermatitidis Culture Identification Test. All non-target isolates representing a phylogenetic cross-section of organisms produced negative results using this kit. The ACCUPROBE Blastomyces Dermatitidis Culture Identification Test does not differentiate between *P. brasiliensis* and *B. dermatitidis* isolates. All *P. brasiliensis* isolates tested in both the filamentous and yeast phases produced positive results. All *B. dermatitidis* isolates tested in both the filamentous and yeast phases produced positive results using the ACCUPROBE Blastomyces Dermatitidis Culture Identification Test.

### D. Recovery

*B. dermatitidis* ribosomal RNA at concentrations ranging from 0.04  $\mu\text{g}$  to 1.6  $\mu\text{g}$  per test was assayed alone and in the presence of 0.0045  $\mu\text{g}$  (equivalent to  $10^5$  cells) and 0.45  $\mu\text{g}$  (equivalent to  $10^7$  cells) ribosomal RNA of either *Histoplasma capsulatum* or *Candida albicans*. There was no interference with the *B. dermatitidis* signal observed and the other organisms present did not react using the ACCUPROBE Blastomyces Dermatitidis Culture Identification Test.

## Bibliography

- Centers for Disease Control and Prevention. 1988. United States Morbid. and Mortal. Weekly Rep. 37:377-382, 387-388.
- Chapman, S. W. 1990. *Blastomyces dermatitidis*, p. 1999. In G. Mandell, R. Douglas, and J. Bennett (ed.), Principles & practices of infectious disease, 3rd ed. Churchill Livingstone, New York.
- Iwen, P. C., L. Sigler, S. Tarantolo, D. A. Sutton, M. G. Rinaldi, R. P. Lackner, D. I. McCarthy, and S. H. Hinrichs. 2000. Pulmonary infection caused by *Gymnascella hyalinospora* in a patient with acute myelogenous leukemia. J. Clin. Microbiol. 38:375-381.
- Khardori, N. 1989. Host-parasite interaction in fungal infections. Eur. J. Clin. Microbiol. Infect. Dis. 8:341-342.
- Kohne, D. E., A. G. Steigerwalt, and D. J. Brenner. 1984. Nucleic acid probe specific for members of the genus Legionella. p. 107-108. In C. Thornsberry, et al. (ed.), Legionella: proceedings of the 2nd international symposium. American Society for Microbiology. Washington, D. C.
- Koneman, E. W., and G. D. Roberts. 1985. Practical laboratory mycology. 3rd ed., p. 109-111. Williams & Wilkins, Baltimore.
- Larone, D. H. 1987. Thermally dimorphic fungi, p. 82-83. In Medically important fungi, a guide to identification. Elsevier Science Publishing Co., Inc., New York.
- McGinnis, M. R. 1980. Laboratory handbook of medical mycology, p. 187-192 and 248-251. Academic Press, New York.
- Restrepo, A. M. 1990. *Paracoccidioides brasiliensis*, p. 2028. In G. Mandell, R. Douglas, and J. Bennett (ed.), Principles & practices of infectious disease, 3rd ed. Churchill Livingstone, New York.
- Rippon, J. W. 1988. The pathogenic fungi and the pathogenic actinomycetes, p. 474-531. In Medical mycology, 3rd ed. W. B. Saunders Company, Philadelphia.
- Shadomy, S., and D. M. Dixon. 1981. *Blastomyces* and *Paracoccidioides*, p. 665. In Medical microbiology and infectious diseases. W. B. Saunders, Philadelphia.
- Shaffer, J. G., and M. Goldin. 1969. Medical mycology; blastomycosis. p. 992-994. In I. Davidsohn and J. B. Henry (ed.), Clinical diagnosis by laboratory methods. W. B. Saunders Company, Philadelphia.

Gen-Probe Incorporated  
10210 Genetic Center Drive  
San Diego, CA 92121

Customer Service: (800) 523-5001  
customerservice@gen-probe.com  
Technical Support: (888) 484-4747  
technicalsupport@gen-probe.com

[www.gen-probe.com](http://www.gen-probe.com)

GEN-PROBE, GEN-PROBE and design, ACCUPROBE, LEADER, and ACCULDR are trademarks of Gen-Probe Incorporated.

U. S. Patent No. 5,558,990, 5,696,251, and 5,723,597.

©1992–2012 Gen-Probe Incorporated  
102963 Rev. K  
2012-01